

## **Materials Research** and Testing

12200 Berlin, Germany Phone: +49 30 8104-0 Fax: +49 30 8112029 Internet: www.bam.de

## Kodak X-Ray Film System

Certificate No: BAM/ZBA/005/06 2<sup>nd</sup> Revised version

Hereby it is confirmed by the BAM Certification Body, that the industrial X-ray film system with the designation

## Film Kodak Industrex AA 400. processed in the developer Kodak Industrex Single Part,

automatic film processing with immersion time of 100 seconds, developer temperature 26 °C

of the manufacturer

Carestream Health, Inc. Rochester, New York, 14608

meets the requirements of system classification C 5 for industrial X-ray films systems for radiographic testing according to EN 584-1 (2006) and ISO 11699-1 (2008) as well as the requirements of system classification II according to ASTM E 1815 (2008).

Certificate holder according to certification contract No. BAM-ZBA-0013-2007-Carestream Health S.A., 1st New Version:

> Carestream Health SA Centre "Les Tuillières" Route de Cité-Ouest, 2 1196 Gland Switzerland

According to standard DIN EN 45 011 (1998) / ISO/IEC Guide 65:1996 the certification comprises a type examination with at least half-yearly periodical surveillance of the products and quarterly surveillance of the manufacturing process data as well as a monitoring of the production site at Windsor, CO (BAM Certification System IIb).

The certificate is valid until 19 December 2015.

BAM test report **BAM VIII.3/6638** of May 9<sup>th</sup> 2006 and the results from current monitoring procedures (reference no. BZS-GS/211/10) are constituent parts of this certificate.

> for BAM Bundesanstalt für Materialforschung und -prüfung Unter den Eichen 87,12205 Berlin, 2010-12-20

Dr. R. Schmidt

**BAM Certification Body** 

Distribution list:: 1st Certificate holder

2nd BAM Certification Body The BAM Certification Body has been accredited according to standard DIN EN 45011:1998 by the DGA mbH (Deutsche Gesellschaft für Akkreditierung mbH). The accreditation is valid for the scope given in certificate DGA-ZE-3998.00.

Dr. U. Ewert

**BAM Assessor** 

Certificates may be published only in full wording and without any additions. Revocable written consent shall be obtained from BAM for changed reproduction and excerpts. The Certificate is issued in English version. In case of dispute the German text is legally binding. Place of jurisdiction is Berlin.

